



SCENAR PROMISES PROGRESS



General SCENAR Concepts: Theory & Practice

Dr.Revenko . Dr.Subbotina -SCENAR Academy, Russia
The educational programs is signed by Russian Ministry of Health

S C E N A R

SCENAR Impulse
BIO-CYBERNETIC SYSTEM “ORGANISM – DEVICE”
POINT ELECTRODE
COMB ELECTRODE
BASIC SETTING
EFFECTS OF SCENAR THERAPY
WE USE SCENAR TO TREAT NEXT DIAGNOSES
SCENAR MECHANISM ACTIONS -
AMPLIFICATION AND CASCADE
SCENAR MECHANISM ACTIONS
BASIS PRINCIPLES OF SCENAR
COMPLAINT “HERE AND NOW”
VECTOR
VECTOR’S TYRES:
GENERAL VECTOR
HORIZONTAL VECTOR
SMALL ASYMMETRY VECTOR
HOW TO WORK IN ZONE
SMALL ASYMMETRY (S As) IN DIAG-0
REDNESS
PALLOR
STICKINESS
SKIN SOUND
PAIN
ALGORHYTHM TO CHOOSE ZONE (no dynamic in local zone)
HORIZONTALS
EXAMPLE FOR NOSE COMPLAINT
HOW TO FIND RECIPROCAL ZONE OF EYE COMPLAINT
RECIPROCATATE ZONES FOR EYE ILLNESSES
EAR MINI-SYSTEM
BODY PROJECTION ON THE FACE
SHEME OF SU – JOCK
SU – JOCK SHEME “INSECT”
GENERAL ZONES
CYCLE
SCENAR treatment stages
DIAG-1 SETTING
SCENAR Impulse
DIGITAL INDEXES IN DIAG-1, 60 HZ
ASSYMETRY IN DIAG-1
PAUSE WHOLISTIC MEDICINE
CONCEPTION OF DISEASE
VETUOLKAS CHEME
HERING RULES
HEALING VECTOR:
GENERAL ZONES- 3 Lines 6 Points

3LINES 6POINTS in DIAG-0
 3 LINES 6 POINTS in DIAG-1
 RULE of 1-st DIFFERENCE +/- 4 - for 3L6P
 RULE of NEXT DOSE +4 - for 3L6P
 RULE of LAST POSITION
 RULE of ZERO
 RULE "Fm VAR"
 RULE "HIGHEST OF ALL"
 FREQUENCY
 IMPULSE NUMBER PER 1 SECOND
 1 IMPULSE IN 1 SECOND- 1 HERZ
 ALGORHYTM TO CHOOSE F
 MODULATION AM
 ALGORHYTHM TO CHOOSE AM
 DEMPHING
 SCENAR Impulse: Damphing
 DEMPHING SETTING AND VAR
 ALGORHYTHM TO CHOOSE DMPH
 INTENCITY
 ALGORHYTHM TO CHOOSE INTENCITY
 ALGORHYTHM TO CHOOSE DELAY - Z
 COLLAR ZONE in DIAG-0
 COLLAR ZONE in DIAG-1
 FULL PROTOCOL «COLLAR ZONE»
 PATIENT'S PALM ZONE
 PALM ZONE in DIAG-1
 ABDOMEN SQUARES
 ABDOMEN SQUARES in DIAG-1
 DIGESTION COMPLAINTS
 CROSS TECHNICS
 NECK RING
 DYNAMIC OF STOMACH ULCERS
 HEART COMPLAINTS
 HEART – PARTICULAR ORGAN – pump, energy, soul
 ANGEL WINGS
 EXTREMITIES COMPLAINTS
 VESSELS ILLNESSES:
 PAIN SYNDROM
 STAIRS WALKING
 ARGUNT SCENAR THERAPY
 ANTI – ALLERGY POINTS
 FIRST AID EFFECT OF SCENAR
 BREATHING COMPLAINTS
 NERVOUS SYSTEM DISORDERS
 RULES OF CRANIO-THERAPY
 SCENAR THERAPY OF FACIAL AND TRIGEMINAL NERVES DISORDERS
 ZONES FOR TREATMENT – NEORISIS, DEPRESSION, PHOBIA
 HOW TO TREAT ADD
 TORTICOLLIS TREATMENT
 DYNAMIC OF SCT OF INSOMNIA
 URINATION AND SEXUAL DISORDERS
 RECIPROCAL ZONES IN VAGINA
 RECOVERING % of prostate complaints, secretion and sexual function
 GYNAECOLOGICAL DISORDERS
 DYNAMIC OF MYOMA
 MENSTRUAL CYCLE DISORDER
 FRIGID, INFERTILITY
 CHILD SCENAR TREATMENT
 IMMUNE POINTS

GENERAL PRINCIPLES TO TREAT ACUTE COMPLAINTS
 GENERAL PRINCIPLES TO TREAT CHRONIC CONDITIONS
 EFFECTS of «TIME VECTORS»
 METHODS
 HOW CONDUCT THE BLANKET PROCEDURE
 PARASYMPATHETIC REACTIONS
 IMMUNE SYSTEM REACTIONS
 HORMONAL SYSTEM REACTIONS
 PSYCHO-EMOTIONAL REACTIONS

#####

3LINES 6POINTS in DIAG-0 –

- 1.Switch on DIAG-0
- 2.Choose settings in accordance with complaints and choose energy A (E)
- 3.Put electrode in vertical position under C-7
- 4.Slowly move the device from C-7 to coccyx, take off and continue sliding from hair to C-7
- 5.Repeat 3 times
- 6.Choose S As and work in 4 vectors until dynamic
- 7.Mark paravertebral part of 3 positions
- 8.Put electrode on the left paravertebral part and move the device down until mark
- 9.Do it the same on the right paravertebral part
- 10.Continue procedure to alternate left and right parts until coccyx, after – do it on the neck
- 11.Repeat 3 times
- 12.Find S As and work in 4 vectors until dynamic
- 13.3 times treat each of 6 Points, find S As and treat until dynamic
-

3 LINES 6 POINTS in DIAG-1

- 1.Setting is: Diag-1, F-60 or 90 Hz, E - comfortable
- 2.Put electrode in vertical position under C-7,note IR, take off the device
- 3.Put electrode on the 2-nd position, measure IR, take off device
- 4.Compare IR-s, if you have difference +/-4, put electrode to the position with highest IR and wait for Dose*
- 5.If you didn't find difference +/-4, move the device to the next position (see slide 158)
- 7.Last position (above C-7)-*
- 8.Do same algorithm on the 2-nd and 3-rd Lines, moving device from left to right
- 9.Last right position - *
- 10.Compare all doses on the 1-st Line, put the device to the position with highest dose (+1) and wait for Zero «0»
- 11.Do the same on the 2-nd and 3-rd Lines – «0»
- 12.If on the 2-nd or 3-rd Line will only one D* - make one “0”
- 13.You can get 2 or 3 “0-s”
- 6.All next positions with IR + 4 should be dosed
- 14. To the highest «0», (+1) – Fm\VAR – 2 min.
- 15. Make rule “Highest of all”
- Reduce E (A) in twice
- Put electrode on the 1-st face point
- Measure IR
- Put the device to the 2-nd point, measure IR
- Find difference + 4* and make Dose*
- Last position / *
- Compare all doses, to the highest + 1\ “0”
- Compare face “0” and spine “0”
- If face “0” + 1 from spine “0”, put Fm\VAR 2 min.
- Make rule “Highest of all”

